

Agilent 4.1
sky 1720
smy 532
coherence sml=341.8 5th harmonic of 1709
DAS profile VSOP_HO
L-band Rx is finally cold
trombone is 110 steps out from IN limit
RX attenuator is 6 dB
Shari did a calibration scan and obtained scale factors of 106.5 and 44.3

From:
<https://www.atnf.csiro.au/vlbi/dokuwiki/> - **ATNF VLBI Wiki**

Permanent link:
<https://www.atnf.csiro.au/vlbi/dokuwiki/doku.php/lbaops/lbajul2009/v289fholog>

Last update: **2015/12/18 16:38**

